





Siemens Medical Solutions USA, Inc. 221 Gregson Drive, Cary, NC 27511

Customer: BRIGHAM AND WOMENS HOSPITAL Customer PO: PO Date:

City: BOSTON Contract#: 0035246625 Contract End date: 31-DEC-2024

State: MA X-Ray Registration Number: N/A

CSE Name			Employee#		fication# Func.Loc.# S		System Description			Room#	
KOPF,JOHN		29955 400112750324 400-289785 SOMATOM Definition A		ATOM Definition AS		L1-CT2					
SVC Arri Date/Time/Day SVC Comp Date/Time/Day		Effect	Cause	Code	S/W Leve	el	Counter		GTIN / Model #	Serial#	
08/07/2024 13:00 Wednesday	08/07/2024 16:05 Wednesday	1	6000						4056869003665	64081	

Symptoms Reported / Reason for Service

No	power	to	gantry.	Table	is	unlocked
----	-------	----	---------	-------	----	----------

Corrective Action / Service Performed

System will not reach standby. found XGS not reading 24vdc. PS is good. Ordered parts as same day.

Repair Hours Breakdown

	Covered hours	Non-Covered hours	
Regular	3.00	0.00	
Overtime	0.00	0.00	
Double Time	0.00	0.00	

Report Sent to following Fax / Email:

none@siemens-healthineers.com ibwh5@bwh.harvard.edu JOHN.KOPF@SIEMENS-HEALTHINEERS.COM

#	Qty	Part # Installed	S/N In	Description	Туре	Qty	Part # Removed	S/N Out
1.								
2.								
3.								
4.								
5.								
6.								
7.								
8.								
9.								
10.								

#	Test Equipment Used	Description	S/N	Calibration Expiration
1.				
2.				
3.				
4.				
5.				

Indicate system was properly functioning and include test & inspection data, if applicable

_				1. 16 c C H	1 24		
S	/stem	requires	a tollow-up	VISIT TO TUILY	/ comply with	manufacturer's	specifications.

This is not an Invoice

Your final invoice will reflect pricing based on your specific Service Agreement coverage and/or Labor Rates including the application of the Flat Travel Fee (FTF).

Customer Signature Date 08/07/2024

(Acknowledgement of Services Rendered)

Dept/Title